Critical Care Intra-Aortic Balloon Pump Guidelines

Critical Care Guidelines FOR CRITICAL CARE USE ONLY

Set-up

Ensure post-insertion ECG, CXR, limb circulation check completed and reviewed

Ensure: Power on

Helium tank open and sufficiently full

Tubing check – connections tight and tubing patent Pressure bag – inflated and transducer zeroed Monitoring – ECG, arterial pressure, balloon pressure

Initial settings: Mode - Auto

Frequency - 1:1 Augmentation – Max

Trigger - ECG

Monitoring

Bedside chart should be completed for every patient

Avoid having pump on standby - risk of clot formation after 10minutes

If the balloon remains on standby for more than 20minutes it should be removed

Continuous monitoring: ECG

Arterial trace

Balloon pressure waveform

Tubing patency

Patient - insertion site, alignment of affected limb

Hourly checks: MAP (from IABP monitor)

Augmented diastolic pressure (from IABP monitor)

frequency

trigger (should be ECG when available) increasing inotrope requirements

bilateral pedal pulse and foot colour/warmth left radial pulse and hand colour/warmth

Daily checks: platelets and U+Es

ECG CXR

Consider iv heparinisation

Weaning

Consider when stable MAP>65, stable heart rate and rhythm, CI >2.2 (if measured)

Process: frequency decreased from 1:1 to 1:2 then 1:4

medical review before and following each reduction augmentation of the balloon should not be altered

Title: Lothian Critical Care Intra-Aortic Balloon Pump Guideline	
	Authors: A Abu-Arafeh G McNeill
Status Draft/Final: Final	Approved by: CC QIT, F Perks Radiology
	Written: August 2018
Reviewed on: Sept 2021	Next review: August 2024


Critical Care Intra-Aortic Balloon Pump Guidelines

Critical Care Guidelines FOR CRITICAL CARE USE ONLY

Reviewing the CXR

Positioning of the balloon tip is important as too proximal a placement risks occlusion of the left subclavian artery while too distal a placement risks renal artery occlusion.

The tip of the balloon pump can be identified on CXR as a small linear metallic marker. During inflation of the balloon, an air filled structure within the descending thoracic aorta may also be visible.



Traditionally the aortic knuckle is used as a CXR landmark, with optimal positioning being 2-4cm below the upper point of the aortic knuckle.



In cardiac failure this is often not easily seen. A position 2cm superior to the carina can also be used to ensure correct position of the balloon tip.

If you have any concerns, please ask a senior colleague to review. A formal radiology report MUST be obtained and reviewed.

Title: Lothian Critical Care Intra-Aortic Balloon Pump Guideline	
	Authors: A Abu-Arafeh G McNeill
Status Draft/Final: Final	Approved by: CC QIT, F Perks Radiology
	Written: August 2018
Reviewed on: Sept 2021	Next review: August 2024

Critical Care Intra-Aortic Balloon Pump Guidelines

Critical Care Guidelines FOR CRITICAL CARE USE ONLY

Troubleshooting

Potential contra-indications to IABP therapy

- Severe aortic regurgitation
- Abdominal or thoracic aortic aneurysm or dissection
- Severe aortic calcification or peripheral vascular disease
- Uncontrolled bleeding disorder or uncontrolled sepsis

Loss of trigger

- check ECG trace and replace ECG electrodes or cable if required
- try alternative ECG lead
- ensure pressure waveform visible and arterial line patent
- check pressure bag inflated and transducer height appropriate

Blood seen in balloon tubing

- STOP pump
- clamp helium line and disconnect balloon tubing from pump
- notify medical team immediately

Increasing inotrope requirements

- do not stop pump
- if no obvious cause (e.g. increased sedation) then notify medical team
- if increased 2 consecutive hours then notify medical team

Absent pulse or abnormal limb colour/warmth

- do not stop pump
- notify medical team immediately

New or sudden onset limb/shoulder pain

- do not stop pump
- notify medical team immediately
- assess limb pulses, colour, temperature

Low helium message or helium tank noted to be low

- no need to stop pump
- pump may need to be restarted if alarm has caused it to pause
- close helium cylinder fully, remove cylinder, check plastic washer is present and not damaged, install new cylinder and slowly open valve

Cardiac arrest

- it is safe to perform DC cardioversion if staff are clear of patient and IABP
- select arterial trigger to continue pump counter-pulsation with CPR effort
- if CPR is not triggering pump then the balloon pump should be switched off and manual inflation considered
 - this involves manually injecting 50ml of air into the balloon and immediately deflating it 5 times every 30minutes. It should only be attempted by staff trained to do so. If the balloon remains deflated for more than 20minutes it should be removed

Title: Lothian Critical Care Intra-Aortic Balloon Pump Guideline	
	Authors: A Abu-Arafeh G McNeill
Status Draft/Final: Final	Approved by: CC QIT, F Perks Radiology
	Written: August 2018
Reviewed on: Sept 2021	Next review: August 2024